



CERTIFICATE OF ACCREDITATION

In terms of section 22(2) (b) of the Accreditation for Conformity Assessment, Calibration and Good Laboratory Practice Act, 2006 (Act 19 of 2006), read with sections 23(1), (2) and (3) of the said Act, I hereby certify that:-

CONCILIUM TECHNOLOGIES (PTY) LTD
Co. Reg. No.: 1999/013330/07
ELECTRICAL DC LOW FREQUENCY LABORATORY

Accreditation Number: **CAL 108-01-00**

is a South African National Accreditation System accredited Calibration Laboratory
provided that all SANAS conditions and requirements are complied with

This certificate is valid as per the scope as stated in the accompanying scope of accreditation
Annexure "A", bearing the above accreditation number for

DC LOW FREQUENCY METROLOGY

The facility is accredited in accordance with the recognised International Standard

ISO/IEC 17025:2017

The accreditation demonstrates technical competency for a defined scope and the operation of a
laboratory quality management system

While this certificate remains valid, the Accredited Facility named above is authorised to use the
relevant SANAS accreditation symbol to issue facility reports and/or certificates

Mr M Phaloane
Acting Chief Executive Officer

Effective Date: 27 February 2024
Certificate Expires: 30 January 2027



ANNEXURE A

SCOPE OF ACCREDITATION

DC LOW FREQUENCY METROLOGY

Accreditation Number: CAL 108-01-00

Permanent Address of Laboratory: Concilium Technologies (Pty) Ltd Electrical DC Low Frequency Laboratory 10 Central Park 13 Esdoring Nook Highveld Technopark Centurion 0157 Postal Address: 10 Central Park 13 Esdoring Nook Highveld Technopark, Centurion 0157 Tel: (012) 678-9200 E-mail: Gavin.Schuster@concilium.co.za	Technical Signatory: Mr GD Schuster Nominated Representative: Mr GD Schuster Issue No.: 18 Date of Issue: 30 August 2024 Expiry Date: 30 January 2027
--	---

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)	METHOD/PROCEDURE
1	DC Voltage (up to 1100 V)			
1.1	DC voltage sources			
1.1.1	Single values			
	1,0 V	DC	$1 \cdot 10^{-6} \cdot U$	Direct comparison with a DC voltage standard.
	10 V	DC	$1 \cdot 10^{-6} \cdot U$	
1.1.2	Low values (≤ 10 V)			
	0 V to 10 V	DC	$1 \cdot 10^{-6} \cdot U + 0,3 \mu V$	Direct measurement or comparison with a DC voltage standard.
1.1.3	Intermediate values (> 10 V to 1 100 V)			
	10 V to 100 V	DC	$3 \cdot 10^{-6} \cdot U$	Direct measurement or comparison with a DC voltage Source or multifunction calibrator.
	100 V to 1 000 V	DC	$5 \cdot 10^{-6} \cdot U$	
1.2	DC voltage meters			
1.2.1	Very low values (≤ 10 V)			
	0 mV to 10 V	DC	$1 \cdot 10^{-6} \cdot U + 0,3 \mu V$	Direct measurement or comparison with a DC Voltage source and short.
1.2.2	Intermediate values (> 1 mV to 1 100 V)			
	10 V to 100 V	DC	$3 \cdot 10^{-6} \cdot U$	Direct measurement or comparison with a DC voltage source or multifunction calibrator.
	100 V to 1 000 V	DC	$5 \cdot 10^{-6} \cdot U$	

Original Date of Accreditation: 1980

Page 1 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 108-01-00
Date of Issue: 30 August 2024
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)	METHOD / PROCEDURE
2	DC Resistance			
2.1	DC Resistance Standards and Sources			
2.1.1	Low values ($\leq 1\Omega$)			
	1 m Ω 10 m Ω 100 m Ω 1 Ω	DC DC DC DC	$2 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-5} \cdot R$	Direct substitution or current volt drop method.
2.1.2	Intermediate values ($>1 \Omega$ to 1 MΩ)			
	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω	DC DC DC DC DC DC	$5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $1 \cdot 10^{-4} \cdot R$	Direct substitution or current volt drop method.
2.1.3	High values ($>1 \text{ M}\Omega$)			
	10 M Ω 100 M Ω 1 G Ω	DC DC DC	$2 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-3} \cdot R$	Direct or substitution method.
2.2	DC Resistance Meters			
2.2.1	Low values ($\leq 1 \Omega$)			
	1 m Ω 10 m Ω 100 m Ω 1 Ω	DC DC DC DC	$2 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-6} \cdot R$	Direct or substitution method.
2.2.2	Intermediate values ($>1 \Omega$ to 1 GΩ)			
	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	DC DC DC DC DC DC DC DC	$5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $5 \cdot 10^{-6} \cdot R$ $5 \cdot 10^{-5} \cdot R$ $1 \cdot 10^{-4} \cdot R$ $2 \cdot 10^{-4} \cdot R$ $5 \cdot 10^{-5} \cdot R$	Direct or substitution method.
2.2.3	High values ($>1 \text{ G}\Omega$)			
	1 G Ω 10 G Ω 100 G Ω	DC DC DC	$1 \cdot 10^{-3} \cdot R$ $2 \cdot 10^{-3} \cdot R$ $5 \cdot 10^{-3} \cdot R$	Direct or substitution method.

Original Date of Accreditation: 1980

Page 2 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 108-01-00
Date of Issue: 30 August 2024
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY OR TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ($\pm$)	METHOD / PROCEDURE
3	DC Current (up to 100 A)			
3.1	DC Current Sources			
3.1.1	Low values ($\leq 0,1$ mA)			
	1 μ A to 100 μ A	DC	$5 \cdot 10^{-5} \cdot I$	Direct or current Volt drop method.
3.1.2	Intermediate values ($> 0,1$ mA to 20 A)			
	0,1 mA to 100 mA 100 mA to 2 A 2 A to 10 A 10 A to 20 A	DC DC DC DC	$5 \cdot 10^{-5} \cdot I$ $2 \cdot 10^{-4} \cdot I$ $2 \cdot 10^{-4} \cdot I$ $5 \cdot 10^{-4} \cdot I$	Direct or current Volt drop method.
3.1.3	High values (> 20 A to 100 A)			
	20 A to 50 A 50 A to 100 A	DC DC	$5 \cdot 10^{-4} \cdot I$ $2 \cdot 10^{-3} \cdot I$	Direct or current Volt drop method.
3.2	Current Meters			
3.2.1	Low values ($< 0,1$ mA)			
	1 μ A to 100 μ A	DC	$5 \cdot 10^{-5} \cdot I$	Direct measurement.
3.2.2	Intermediate values (0,1 mA to 20 A)			
	0,1 mA to 100 mA 100 mA to 2 A 2 A to 10 A 10 A to 20 A	DC DC DC DC	$5 \cdot 10^{-5} \cdot I$ $2 \cdot 10^{-4} \cdot I$ $2 \cdot 10^{-4} \cdot I$ $2 \cdot 10^{-3} \cdot I$	Direct measurement.
3.2.3	High values (> 20 A to 100 A)			
	20 A to 50 A	DC	$2 \cdot 10^{-3} \cdot I$	Direct measurement.
4	Impedance (up to the MHz range)			
4.2	Capacitance			
4.2.1	Standard Capacitors			
	100 pF to 1 μ F 1 μ F to 10 μ F 10 μ F to 100 μ F	120 Hz to 10 kHz 120 Hz to 10 kHz 120 Hz to 1 kHz	$1 \cdot 10^{-3} \cdot C$ $2 \cdot 10^{-3} \cdot C$ $2 \cdot 10^{-3} \cdot C$	Direct measurement or substitution method.
4.2.2	Fixed Capacitor			
	100 pF to 1 μ F 1 μ F to 10 μ F 10 μ F to 100 μ F	120 Hz to 10 kHz 120 Hz to 10 kHz 120 Hz to 1 kHz	$1 \cdot 10^{-3} \cdot C$ $2 \cdot 10^{-3} \cdot C$ $2 \cdot 10^{-3} \cdot C$	Direct measurement or substitution method.
4.2.4	Capacitance Meters and Bridges			
	1 pF, 10 pF, 100 pF and 1 nF	50 Hz to 1 kHz 1 kHz to 1 MHz	$5 \cdot 10^{-5} \cdot C$ $5 \cdot 10^{-4} \cdot C$	Direct measurement of reference standard capacitor.
	10 nF	50 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz	$5 \cdot 10^{-5} \cdot C$ $5 \cdot 10^{-4} \cdot C$ $1 \cdot 10^{-3} \cdot C$ $1 \cdot 10^{-2} \cdot C$	
	100 nF	50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 200 kHz	$5 \cdot 10^{-5} \cdot C$ $5 \cdot 10^{-4} \cdot C$ $3 \cdot 10^{-3} \cdot C$	
	1 μ F	50 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz	$5 \cdot 10^{-5} \cdot C$ $1 \cdot 10^{-3} \cdot C$ $2 \cdot 10^{-3} \cdot C$	
	10 μ F and 100 μ F	120 Hz to 1 kHz	$2 \cdot 10^{-3} \cdot C$	

Original Date of Accreditation: 1980

Page 3 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

Accreditation Manager

ANNEXURE A

Accreditation No.: CAL 108-01-00
Date of Issue: 30 August 2024
Expiry Date: 30 January 2027

ITEM	MEASURED QUANTITY, TYPE OF GAUGE OR INSTRUMENT AND RANGE OF MEASURED QUANTITY	NOMINAL FREQUENCY	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	METHOD / PROCEDURE
4 4.3 4.3.2	Impedance (up to the MHz range)			
	Inductance			
	Fixed Inductor, Variable Inductor box (1 mH to 1 H)			
	1 mH to 1 H	120 Hz to 1 kHz	$2 \cdot 10^{-3} \cdot L$	Direct measurement or substitution method.
	Fixed Inductor, Variable Inductor box (> 1 H)			
	1 H to 10 H	120 Hz	$5 \cdot 10^{-3} \cdot L$	Direct measurement or substitution method.
	4.3.5	Inductance Meters and Bridges		
1 mH to 1 H		120 Hz to 1 kHz	$2 \cdot 10^{-3} \cdot L$	Direct measurement.
1 H to 10 H		120 Hz	$5 \cdot 10^{-3} \cdot L$	
5 5.2.1	AC Voltage			
	AC Voltage Sources (up to 1 000 V)			
	0 V to 10 V	30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz	$1 \cdot 10^{-4} \cdot U + 10 \mu V$ $2 \cdot 10^{-4} \cdot U + 10 \mu V$ $3 \cdot 10^{-4} \cdot U + 10 \mu V$	Direct measurement.
	10 V to 20 V	30 Hz to 20 kHz 20 kHz to 100 kHz	$1 \cdot 10^{-4} \cdot U$ $2 \cdot 10^{-4} \cdot U$	
	20 V to 200 V	30 Hz to 20 kHz 20 kHz to 100 kHz	$1 \cdot 10^{-4} \cdot U$ $2 \cdot 10^{-4} \cdot U$	
	200 V to 1 000 V	30 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	$3 \cdot 10^{-4} \cdot U$ $4 \cdot 10^{-4} \cdot U$ $6 \cdot 10^{-4} \cdot U$	
	5.2 5.2.2	AC Voltage (up to 1 000 V)		
Meters				
0 V to 10 V		30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz	$1 \cdot 10^{-4} \cdot U + 10 \mu V$ $2 \cdot 10^{-4} \cdot U + 10 \mu V$ $3 \cdot 10^{-4} \cdot U + 10 \mu V$	Direct measurement.
10 V to 20 V		30 Hz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz	$1 \cdot 10^{-4} \cdot U$ $2 \cdot 10^{-4} \cdot U$ $3 \cdot 10^{-4} \cdot U$ $4 \cdot 10^{-4} \cdot U$	
20 V to 200 V		30 Hz to 20 kHz 20 kHz to 100 kHz	$1 \cdot 10^{-4} \cdot U$ $2 \cdot 10^{-4} \cdot U$	
200 V to 1 000 V		50 Hz to 1 kHz	$3 \cdot 10^{-4} \cdot U$	
6 6.2	AC current			
	AC current (up to 100 A)			
6.2.1	Sources			
	0 A to 2 A	30 Hz to 5 kHz	$3 \cdot 10^{-4} \cdot I + 2 \mu A$	Direct measurement or volt drop method.
6.2.2	Meter			
	0 A to 2 A	30 Hz to 5 kHz	$3 \cdot 10^{-4} \cdot I + 2 \mu A$	Direct measurement or volt drop method.

Original date of accreditation: 1980

Page 4 of 4

The CMC, expressed as an expanded uncertainty of measurement, is stated as the standard uncertainty of measurement multiplied by a coverage factor $k = 2$, corresponding to a confidence level of approximately 95%

ISSUED BY THE SOUTH AFRICAN NATIONAL ACCREDITATION SYSTEM

Accreditation Manager